

Long-Lasting Aqueous Mountants for Light Microscopy

John A. Kiernan,

The University of Western Ontario

The requirements for such a medium are that it won't deposit crystals when it evaporates and that it won't be incapable of evaporating. The second condition probably excludes anything containing glycerol. The first is more difficult. A polymer such as gum, polyvinylpyrrolidone or gelatin might fill the bill, especially in the presence of a sugar that can form a syrup that's almost solid. Fructose (levulose) has this reputation, and is a good aqueous mountant. Very cheap, but for some reason somewhat too acidic for basic dyes, but can be buffered. It's not 100% predictable. I've had fructose crystallize under coverslips, but you can quickly remove all the mess with water and then try again.

An old and certain way to make a solid and perhaps permanent aqueous mount is to use a smaller-than-usual coverslip and as little aqueous mountant as possible. Allow a few days for water evaporation at the edges of the coverslip, otherwise the refractive index will be too low and unstained parts of the section will be visible. Sometimes that is wanted, in which case just mount in a minimal drop of glycerol under the small coverslip. It's important to keep the top of the coverslip completely free of dust, condensed moisture and anything else insoluble in xylene. When the aqueous mount with its little coverslip looks OK, give it a swish in clean xylene to remove any grease and apply a larger coverslip in the usual way, as if the preparation were a very thick (200 μm) section. The aqueous mount is now sealed into a resinous medium.

The problem with this is that coverslip thickness is at least 340 μm . This reduces the resolution of a microscope objective without an adjustable "coverslip collar," and if the section itself is a thick one, its lower levels may be beyond the working distance of anything cheaper than a \$20,000 water-immersion objective. This won't be a problem unless you're counting mitochondria or peroxisomes, but the unduly thick double coverslip can still be annoying, because it necessitates much fiddling with the fine focus and substage aperture. These annoyances can be reduced.

A refinement of the above technique is especially valuable for cryostat sections collected onto coverslips rather than slides. Use a small or a large coverslip to collect the section. After staining, make the aqueous mount on a larger or smaller coverslip. Wait for evaporation, etc., as explained above, and then do the resinous mounting, with the larger coverslip uppermost and the smaller one applied to the slide. Now there is only the thickness of one coverslip (and an unknown but small depth of one mounting medium) between the specimen and the front lens of the objective.

For dehydrogenase histochemistry (seeing mitochondria in unfixed cells, usually in tissue whose general microanatomy suffers quite a lot to get this result), collection of sections on coverslips is a must. For lipid staining with Sudan dyes, it's often OK to collect the sections on slides and have double thickness coverslipping, because high resolution may not be needed. Never use any product without finding out and learning about its composition. The results of research are questionable if any part of the methodology is not fully understood by the investigator. This applies even to a mounting medium. How certain can you be that it faithfully preserves the stain? And for how long? And how do you know it's that good? ■

Front Page Image Outer Surface of Egg Shell

Landscape Format – Original Mag.: 1,500X – Final Mag.: 3,100X – (10 μm = 31 mm)

The image was taken on a JEOL JSM-840 SEM at 5 kV. Digital files were captured at a resolution of 1280 x 960 with a 160 second dwell time. The image was manipulated as indicated in the article "Making Anaglyph Images from SEM Images Using Adobe PhotoShop" by Debby Sherman on page 8 of this issue, and was printed on a Codonics 1660 dye-sublimation printer. Stereo pairs were separated by a 7° tilt.

★ A STEREO IMAGE – VIEW WITH RGB GLASSES ★

MICROSCOPY TODAY

The objective of the publication, perhaps unlike many others, is to present articles and other material of interest and value to the working microscopist. With contributions from our readership, we attempt to cover all aspects of microscopy. The publication is mailed, ten times a year, at no charge to some 8,200 microscopists in the United States - all of which have requested subscriptions. Due to the current relatively low number of international readers, and resulting very high postage costs, we are forced to charge the following for international subscriptions (10 issues/year):

Canada, Mexico, Latin and South America: \$40 for one year, \$70 for two years

Other International: \$55 for one year, \$95 for two years

Payment may be made by check, credit card or company purchase order.


MICROSCOPY TODAY

PO Box 620122, Middleton, WI 53562

Tel.: (608)836-1970 - Fax: (608)836-1969 - eMail: MicroToday@aol.com - <http://www.microscopy-today.com>

Copyright in 1998 by Today Enterprises. All rights reserved.

Incredible



for \$80,000

A Complete SEM/EDX System Without Any Compromises

Incredible to say the least. A really exceptional value. It's a *fully integrated system* of the versatile, easy-to-use Personal SEM with the trusted Personal EDX. It's truly integrated – *PC controlled* by a single graphical user interface (GUI). Plus, *full digital imaging*. For comprehensive analysis, you have X-ray Line, spot and dot mapping. There are no compromises here.

Instead of a diffusion pump, we incorporate a *turbo pump* for a cleaner vacuum, and you'll find it easier to work with. Looking to go down to lighter elements? Well, our system includes a *light element detector* for detection of elements down to Boron. Concerned about maintenance? Don't worry. Our SEM/EDX has a *disposable column liner*

and emitter for easy maintenance. In fact, don't worry about the entire system. There's a *one year warranty*.

Looking to the future, R J Lee Instruments is a *single source supplier* – service and support for both the SEM and EDX are from the same source. As your needs change, our SEM/EDX system is both *fully upgradeable and fully expandable*.

When you examine all the features and attributes of the R J Lee SEM/EDX, you'll see that this is an *incredible value at \$80,000 and there aren't any compromises*.

For more information or to have a representative call you, visit our web site rjleeinst.com or call 724-744-0100.

Microanalysis

THE APPLICATIONS SPECIALISTS

R J Lee Instruments Limited

Circle Reader Inquiry #10