

TABLE

COMPARISON OF MATCHED-PAIR RESPONSES ON PRE- AND POST-USE QUESTIONNAIRES (N=17)*

Item	Median Pre-Use Score	Median Post-Use Score	P†
Overall perception			
and satisfaction	8	5	.000
Ease of use	8	5	.001
Ease of filling	6	5	.019
Accuracy of delivery	6	4.5	.011
Location of dose			
markings	6.5	3.5	.003
Ease of locking			
and unlocking	7	4	.002
Patient safety and			
comfort	6	5	.695
Personal safety	10	8	.012

*Comparison of pre-use and post-use responses to the questions, "Compared to the syringes you normally use, how would you rate the safety syringe?" Answers were selected from a 10-point Likert scale, with 1=poor, 5=average, and 10=excellent.

† Wilcoxon Signed Rank Test.

uated engineering controls for several devices.³ Of the 390 syringes collected from the needle disposal boxes during our study period, only 59.5% were activated.

Our study of this new safety syringe had several limitations, including a small sample size, such that the expected number of needlestick injuries during the study period was nil. Thus, we did not expect to measure an actual decrease in the number of needlestick injuries as a result of the new syringe. Underreporting of needlesticks is common in the medical field as a whole,⁴⁻⁶ and so it would be naive to think that it did not affect this study as well. It is possible that a needlestick injury did occur with the safety syringe and never was reported.

We cannot clearly establish that the safety syringes in fact were activated immediately after use. It is possible the safety device was not activated immediately after use, but only upon disposal, which would not protect the HCW and other staff from accidental needlestick injuries while the uncapped needle was transported to a distant disposal site or was put down on the bed or adjacent table while the patient was stabilized.

Finally, this study relied on self-reported data, which may not be entirely accurate.

We have demonstrated that the correct use of this safety syringe was at most 60%, even after an educational presentation to a group of experienced nursing personnel. The weakest link in any active safety device clearly is the need for the user to activate the safety feature through one or more additional steps that take time, effort, and remembering to perform the task each time. Based on our results, we would suggest that passive rather than active safety products be developed.

95-CC-039. Address reprint requests to Leland S. Rickman, MD, Epidemiology Unit, UCSD Medical Center 8951, 200 W Arbor Dr, San Diego, CA 92103.

REFERENCES

1. Jagger J, Hunt E, Pearson RD. Estimated cost of needlestick injuries for six major needled devices. *Infect Control Hosp Epidemiol* 1990;11:584-588.
2. Haiduvon DJ, Stevens DA. Eight-year analysis of percutaneous injuries (PI): categorization, effective reduction methods and future strategies. Presented at the 34th Interscience Conference on Antimicrobial Agents and Chemotherapy; October 6, 1994; Orlando, FL. Abstract J141.
3. Short L, Chamberland M, Gerberding J, et al. Implementation of engineering controls (ECs) to reduce percutaneous injuries during phlebotomy (PIPs). Presented at the 34th Interscience Conference on Antimicrobial Agents and Chemotherapy; October 6, 1994; Orlando, FL. Abstract J139.
4. Ferguson TJ. Needle-stick injuries among health care professionals. *West J Med* 1992;156:409.
5. Tandberg D, Stewart KK, Doezeema D. Under-reporting of contaminated needlestick injuries in emergency health care workers. *Ann Emerg Med* 1991;20:66-70.
6. Morgan DR. HIV and needlestick injuries. *Lancet* 1990;335:1280. Letter.

HIV Origins

Gina Pugliese, RN, MS
Martin S. Favero, PhD
Medical News Editors

The link between human immunodeficiency virus (HIV) and simian immunodeficiency virus (SIV) has been debated for years. In a recent article in *Lancet*, Dr. Francoise Barre-Sinoussi, of the Pasteur Institute in Paris, suggests that SIV may be a valuable model for studying HIV and AIDS. Human immunodeficiency virus shows extensive genetic diversity, even within patients. Evolution of viral isolates increases as the clinical disease progresses, allowing HIV-1 to

develop resistance to antiviral drugs quickly. Genetic analysis has found two distinct groups of HIV-1 isolates—the M group and the O group. The M group, which contains most HIV-1 isolates, is subdivided into at least 10 strains.

The discovery of HIV-1 in West Africa and its similarity to SIV has raised the possibility of a link between human and nonhuman primate lentiviruses. The divergence between HIV-1 group M subtypes and SIV suggests that chimpanzee viruses may have been introduced into the human population 30 to 50 years ago. The origin of HIV-1 group O remains to be

explained, however. While genetically similar to HIVs, SIVs do not cause disease in their natural hosts, although cross-species transmission may result in pathogenic infections. Barre-Sinoussi concludes that nonhuman lentiviruses from Old World primates can provide useful models for studying the interaction between the human host and viral factors that cause AIDS.

FROM: Barre-Sinoussi F. HIV as the Cause of AIDS. *Lancet* 1996; 348:31-32.

National AIDS Clearinghouse. AIDS Daily Summary, Bethesda, MD, July 11, 1996.